

UL TEST REPORT AND PROCEDURE

Standard:	ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)(Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance) CAN/CSA-C22.2 No. 60601-1 (2008) (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance)
Certification Type:	Component Recognition
CCN:	QQHM2, QQHM8 (Power Supplies, Medical and Dental)
Product:	AC/DC Medical Adapter
Model:	MDS-150AAS12 B MDS-150AAS19 B MDS-150AAS24 B
Rating:	Input rating: 100 ~ 240Vac; 2.5-1A; 50-60 Hz Output rating: MDS-150AAS12 B 12Vdc, 10A MDS-150AAS19 B 19Vdc, 7.89A MDS-150AAS24 B 24Vdc, 6.25A
Applicant Name and Address:	DELTA ELECTRONICS INC 3 TUNGYUAN RD CHUNGLI INDUSTRIAL ZONE TAOYUAN COUNTY 32063 TAIWAN

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL LLC ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.

Prepared by: Christine, Wang

Reviewed by: Jimmy Deng