



Underwriters Laboratories (UL LLC) Safety Certification Report

Model: IMA-X400-Y-ZZZZZ (where X = C or S, Y = 12 thru 48, Z = 0-9 or A-Z or blank)

Device Description: Switching Power Supply for Building In

Applicant: DELTA ELECTRONICS (THAILAND) PUBLIC CO LTD
909 SOI 9, MOO 4,PATTANA 1RD
BANGPOO INDUSTRIAL ESTATE (E P Z)
TAMBOL PHRAKSA AMPHUR MUANG
SAMUTPRAKARN 10280 THAILAND

Manufacturer: Same as Applicant

Manufacturing Facility(ies): Same as Applicant

Report No.: E325662-D1011-1/A0/C0-UL-UL

Report (Re)Issue Date: 2017-12-19

Base Standard(s): ANSI/AAMI ES60601-1:2005/(R)2012 and A1:2012, C1:2009/(R)2012 and A2:2010/(R)2012, CAN/CSA C22.2 No. 60601-1:14

Additional Standards: EN 60601-1:2006/A1:2013/A12:2014

Report Types: This report consists of the following report types:
[Yes] US Certification (UL Recognition)
[Yes] CAN Certification (cUL Recognition)

This report covers the Safety evaluation of the referenced model(s) according to the standard(s) specified above.